



Certificate of Compliance

Certificate: 2494607

Master Contract: 163467

Project: 2494607

Date Issued: April 18, 2012

Issued to: Spellman High Voltage Electronics
Corporation
475 Wireless Blvd
Hauppauge, NY 11788
USA
Attention: Godrej Mehta

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Manoj Waghela

Issued by: Manoj Waghela

PRODUCTS

CLASS 8780 01 - MEDICAL ELECTRICAL EQUIPMENT/SYSTEMS

CLASS 8780 81 - MEDICAL ELECTRICAL EQUIPMENT/SYSTEMS - Certified to US Standards

High Voltage X-Ray Generator: Model CRX667LX3950 permanently installed rated: 3 phase 480 Vac, 50/60 Hz, 80A pulsed, 120Vac, Single phase 50/60Hz, 4A. Consists of following:

Cathode Power Module: (P/N 405794-021) -75kV Max, 667mA Max, 40kW max

Filament: 15V @ 8A max.

Grid (G1, G2) -3000Vmax, 4kHz max, 300pfd load max

Anode Power Module: (P/N 405794-022) +75kV Max, 667mA Max, 40kW Max.

HSS: 190V @ 11A 5 sec max.

280V @ 6.2A, 40Sec max

210V @ 4A continuous.