

CERTIFICATE OF COMPLIANCE

Certificate Number 20151120-E227588
Report Reference E227588-A29-UL
Issue Date 2015-NOVEMBER-20

Issued to: SPELLMAN HIGH VOLTAGE ELECTRONICS CORP
475 WIRELESS BLVD
HAUPPAUGE NY 11788

This is to certify that representative samples of COMPONENT - POWER SUPPLIES FOR MEASUREMENT, CONTROL AND LABORATORY USE
See Addendum

Have been investigated by UL in accordance with the Standard(s) indicated on this Certificate.

Standard(s) for Safety: UL 61010-1 & CAN/CSA-C22.2 No. 61010-1, ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - Part 1: General Requirements

Additional Information: See the UL Online Certifications Directory at www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's Certification and Follow-Up Service.

Recognized components are incomplete in certain constructional features or restricted in performance capabilities and are intended for use as components of complete equipment submitted for investigation rather than for direct separate installation in the field. The final acceptance of the component is dependent upon its installation and use in complete equipment submitted to UL LLC.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>



CERTIFICATE OF COMPLIANCE

Certificate Number 20151120-E227588
Report Reference E227588-A29-UL
Issue Date 2015-NOVEMBER-20

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

High Voltage Power Supply –
uX Series: uX50P50, uX50P75 and uX65P65:

where the model number consists of 'uX' or 'uXHP' followed by 50 thru 65 indicating the maximum output voltage in kV'; followed by either the letter 'P' for positive or 'N' for negative, indicating the HV Output polarity; followed by '50', '65' or '75' respectively, indicating the peak power in watts; may be followed by GB or XCC, may also be followed by the letter 'X' and four digits which indicate non-safety related customer options.

uXHP Series: uXHP50P100, uXHP65P100, uXHP80P100:

where the model number consists of 'uXHP' followed by '50' or '65' or '80' indicating the maximum output voltage in kV'; followed by either the letter 'P' for positive or 'N' for negative, indicating the HV Output polarity; followed by '100', indicating the peak power in watts; may be followed by GB or XCC, may also be followed by the letter 'X' and four digits which indicate non-safety related customer options.



Bruce Mahrenholz, Director North American Certification Program
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>

