

CERTIFICATE OF COMPLIANCE

Certificate Number 20120709-E354595
Report Reference E354595-A2-UL
Issue Date 2012-JULY-09

Issued to: SPELLMAN HIGH VOLTAGE LTD
UNIT 14, BROOMERS PARK
BROOMERS HILL LANE, PULBOROUGH, WEST SUSSEX
RH20 2RY UNITED KINGDOM



**This is to certify that
representative samples of** COMPONENT - POWER SUPPLIES FOR MEASUREMENT,
CONTROL AND LABORATORY USE
See Addendum Page

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety: Standard for Electrical Equipment for Measurement, Control, and
Laboratory Use; Part 1: General Requirements, UL 61010-1;
Standard for Electrical Equipment for Measurement, Control, and
Laboratory Use; Part 1: General Requirements, CAN/CSA-C22.2
No. 61010-1.

Additional Information: See the UL Online Certifications Directory at
www.ul.com/database for additional information

Only those products bearing the UL Recognized Component Marks for the U.S. and Canada should be
considered as being covered by UL's Recognition and Follow-Up Service and meeting the appropriate U.S. and
Canadian requirements.

The UL Recognized Component Mark for the U.S. generally consists of the manufacturer's identification and
catalog number, model number or other product designation as specified under "Marking" for the particular
Recognition as published in the appropriate UL Directory. As a supplementary means of identifying products that
have been produced under UL's Component Recognition Program, UL's Recognized Component Mark: , may
be used in conjunction with the required Recognized Marks. The Recognized Component Mark is required when
specified in the UL Directory preceding the recognitions or under "Markings" for the individual recognitions. The
UL Recognized Component Mark for Canada consists of the UL Recognized Mark for Canada:  and the
manufacturer's identification and catalog number, model number or other product designation as specified under
"Marking" for the particular Recognition as published in the appropriate UL Directory.

The final acceptance of the component is dependent upon its installation and use in complete equipment
submitted to UL LLC.

Look for the UL Recognized Component Mark on the product.

William R. Carney, Director, North American Certification Programs
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please
contact a local UL Customer Service Representative at www.ul.com/contactus



CERTIFICATE OF COMPLIANCE

Certificate Number 20120709-E354595
Report Reference E354595-A2-UL
Issue Date 2012-JULY-09

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

High Voltage Power Supply, Models: MPS30P10/24, MPS30N10/24:

where:

- The number 30 = output voltage in kV,
- P/N indicates positive or negative output.
- The number 24 indicates the input voltage.

William R. Carney, Director, North American Certification Programs

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at www.ul.com/contactus

