

CERTIFICATE OF COMPLIANCE

Certificate Number 20130829-E354595
Report Reference E354595-A11-UL
Issue Date 2013-AUGUST-29



Issued to: SPELLMAN HIGH VOLTAGE LTD
UNIT 14, BROOMERS PARK
BROOMERS HILL LANE
PULBOROUGH
WEST SUSSEX
RH20 2RY UNITED KINGDOM

This is to certify that representative samples of COMPONENT - POWER SUPPLIES FOR MEASUREMENT, CONTROL AND LABORATORY USE
See Addendum Page

Have been investigated by UL in accordance with the Standard(s) indicated on this Certificate.


Standard(s) for Safety: See Addendum Page
Additional Information: See the UL Online Certifications Directory at www.ul.com/database for additional information

Only those products bearing the UL Recognized Component Marks for the U.S. and Canada should be considered as being covered by UL's Recognition and Follow-Up Service and meeting the appropriate U.S. and Canadian requirements.

The UL Recognized Component Mark for the U.S. generally consists of the manufacturer's identification and catalog number, model number or other product designation as specified under "Marking" for the particular Recognition as published in the appropriate UL Directory. As a supplementary means of identifying products that have been produced under UL's Component Recognition Program, UL's Recognized Component Mark: , may be used in conjunction with the required Recognized Marks. The Recognized Component Mark is required when specified in the UL Directory preceding the recognitions or under "Markings" for the individual recognitions. The UL Recognized Component Mark for Canada consists of the UL Recognized Mark for Canada:  and the manufacturer's identification and catalog number, model number or other product designation as specified under "Marking" for the particular Recognition as published in the appropriate UL Directory.

Recognized components are incomplete in certain constructional features or restricted in performance capabilities and are intended for use as components of complete equipment submitted for investigation rather than for direct separate installation in the field. The final acceptance of the component is dependent upon its installation and use in complete equipment submitted to UL LLC.

Look for the UL Recognized Component Mark on the product.



William R. Carney, Director, North American Certification Programs

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at www.ul.com/contactus



CERTIFICATE OF COMPLIANCE

Certificate Number 20130829-E354595
Report Reference E354595-A11-UL
Issue Date 2013-AUGUST-29

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

High Voltage Power Supply :

MTS50P50/24/XXX or MTS50N50/24/XXX

Where :

The number 50 after the MTS prefix = output voltage in kV

P/N indicates positive / negative output (selected at manufacture)

50 indicates the output power in watts

24 indicates the input voltage


The XXX number = 616, 617, 738 or 810 is used to specify a specific customer option

Standard(s) for Safety

UL 61010-1 : ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - Part 1:

General Requirements

CAN/CSA-C22.2 No. 61010-1 : ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - Part 1: General Requirements



William R. Carney, Director, North American Certification Programs

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at www.ul.com/contactus

